

## **A novel method of using accelerometry for upper limb FES control.**

SUN, Mingxu, KENNEY, Laurence, SMITH, Christine <<http://orcid.org/0000-0001-5354-953X>>, WARING, Karen, LUCKIE, Helen, LIU, Anmin and HOWARD, David

Available from Sheffield Hallam University Research Archive (SHURA) at:

<https://shura.shu.ac.uk/13946/>

---

This document is the Accepted Version [AM]

### **Citation:**

SUN, Mingxu, KENNEY, Laurence, SMITH, Christine, WARING, Karen, LUCKIE, Helen, LIU, Anmin and HOWARD, David (2016). A novel method of using accelerometry for upper limb FES control. Medical engineering & physics, 38 (11), 1244-1250. [Article]

---

### **Copyright and re-use policy**

See <http://shura.shu.ac.uk/information.html>

# **A novel method of using accelerometry for upper limb FES control**

Mingxu Sun<sup>1</sup>, Laurence Kenney<sup>1\*</sup>, Christine Smith<sup>2</sup>, Karen Waring<sup>1</sup>, Helen Luckie<sup>1</sup>, Anmin Liu<sup>1</sup>, David Howard<sup>1,3</sup>

<sup>1</sup>Centre for Health Sciences Research, University of Salford, UK

<sup>2</sup>Department of Allied Health Professions, Sheffield Hallam University, UK

<sup>3</sup>School of Computing, Science and Engineering, University of Salford, UK

\* Corresponding author.

Tel.: +44 161 295 2289

Email: [l.p.j.kenney@salford.ac.uk](mailto:l.p.j.kenney@salford.ac.uk)

## **Abstract**

This paper reports on a novel approach to using a 3-axis accelerometer to capture body segment angle for upper limb functional electrical stimulation (FES) control. The approach calculates the angle between the accelerometer x-axis and the gravity vector, while avoiding poor sensitivity at certain angles and minimising errors when true acceleration is relatively large in comparison to gravity. This approach was incorporated into a state-machine controller which is used for the real-time control of FES during upper limb functional task performance. An experimental approach was used to validate the new method. Two participants with different upper limb impairments resulting from a stroke carried out four different FES-assisted tasks. Comparisons were made between angle calculated from arm-mounted accelerometer data using our algorithm and angle calculated from limb-mounted reflective marker data. After removal of coordinate misalignment error, mean error across tasks and subjects ranged between 1.4 and 2.9 degrees. The approach shows promise for use in the control of upper limb FES and other human movement applications where true acceleration is relatively small in comparison with gravity.

*Keywords: Accelerometer, Body segment angle, Functional electrical stimulation control, Upper limb rehabilitation, Functional tasks*

## **Introduction**

A recent systematic review concluded that the use of functional electrical stimulation (FES) to promote recovery of upper limb activity after stroke was significantly more effective than activity training alone[1]. Although the finding was positive, at least 4 of the included studies were based on systems which either did not encourage voluntary effort to achieve a functional task, and/or did not offer flexibility over the sequence of stimulation delivered [2-5]. Both voluntary engagement with, and variation in, task practice are considered to be important elements of rehabilitation

programmes for motor re-learning [6, 7], suggesting limitations with the technologies used in these studies.

In recognition of this, researchers have developed a number of systems which provide the patient with FES support for a range of tasks and encourage voluntary engagement of the user. These include devices based on iterative learning control [8], proportionally controlled systems using EMG as an input signal [9] and, of most relevance to this paper, user-defined state machine controlled systems [10, 11]. User-defined state machine control offers a potentially simple approach to implementing task- and patient-specific FES support. However, in order to provide the user with the opportunity to directly engage with the task by actively initiating or controlling stimulation, a method is required to detect user intent. A range of sensors for this purpose have been investigated, including EEG [12] and EMG[13]. Each of these sensing approaches has its limitations and for all but the most severely impaired, limb-mounted movement sensors offer an attractive alternative. Although 6 degree of freedom inertial measurement units (IMUs), which typically include accelerometers, rate gyroscopes and magnetometers, are available, they remain relatively bulky. This precludes their use on, for example, individual fingers. Further, although the price of IMUs has reduced over the past few years, compared with these devices, accelerometers remain cheaper, simpler to use in terms of data processing, and lower power. A small number of researchers have therefore investigated their application to upper limb FES [14-17]. However, as will be discussed below the methods used to derive estimations of orientation from accelerometer data are limited in these and many other papers.

Current methods for processing accelerometer signals to estimate orientation can be categorised according to the number of independent measures (accelerometer axes) used. The first group of methods are based on using just one accelerometer signal, either a single axis device or one axis of a multi-axis device. In the first embodiment of these, a series of calibration measurements over a range of angles is required and an interpolation algorithm can then be used to derive the angle of the sensitive axis from the vertical [18-20]. As the calibration curve is significantly non-linear, the accuracy is highly dependent on a thorough calibration. More commonly, a trigonometric approach is used, based on arccos or arcsin functions (e.g. [21-24]). Both of these approaches require calibration to accurately identify the value of the denominator. Regardless of which of these three techniques is adopted for processing the accelerometer signal, they all suffer from the same drawbacks. When the magnitude of acceleration on the sensitive axis approaches either 9.81 or -9.81, the sensitivity approaches zero because  $\sin \beta$  or  $\cos \beta$  approach 1, which means the signal to noise ratio is very poor [21, 23, 25, 26]. A small number of papers ignore the issue of poor sensitivity at zero or 90 degrees by suggesting a workable range for measurements [21, 23, 24]. For example, Miroslav Husak [23] refers to the measurement range with zero sensitivity as being a “Dead zone” and reports that, using the arcsin function, error increases from less than 2° over 4° when the sensitive axis nears  $\pm 90^\circ$ .

The second group of methods uses a dual axis accelerometer (or two axes of a 3-axis device). The signals from both of the sensitive axes can be used to calculate the angle from the vertical  $\theta = \arctan\left(\underline{g}_z/\underline{g}_x\right)$ . This method suffers from decreasing sensitivity and, hence, increasing angle errors as  $\theta$  approaches 0° [27-29] and extreme sensitivity near  $\pm 90^\circ$  [27, 28]. For example, Šipoš et

al [29] reports that, using the arctan function, error increases from  $1.37^\circ$  to over  $4.5^\circ$  when the sensitive axis nears  $0^\circ$ .

When applying either approach to the analysis of accelerometers signals there remains the problem that orientation can only be accurately estimated if the true acceleration is small compared with gravity (as accelerometers measure the sum of true acceleration and gravity). While the use of distributed multiple accelerometers provide a solution to this problem [30, 31], this is less than ideal for application to upper limb FES because of the difficulty of donning the additional, rigidly connected, accelerometers.

In upper limb FES applications, the way the accelerometer signals are used for control purposes can be categorised as either direct use of the raw accelerometer signals [14, 16, 17] or using an angle-based approach [15], both of which suffer from one or more of the problems listed above.

In this paper we report on a novel approach to using accelerometry for upper limb FES control which addresses the issues discussed above. The paper begins by introducing a new approach to estimating angle relative to gravity from a 3D accelerometer, based on vector mechanics, which includes methods for avoiding poor sensitivity at certain angles and minimising errors when true acceleration is relatively large in comparison to gravity. The paper then reports on its application in our finite-state-machine (FSM) controlled FES system and evaluates its performance.

## **Methods**

### ***Angle estimation algorithm***

The new method calculates the absolute angle from the vertical of one axis of a 3-axis accelerometer (for the purpose of the paper, the x-axis). Referring to Figure 1, the method calculates the angle  $\beta$  between  $\hat{x}$  and  $\underline{g}_{now}$ , where  $\hat{x}$  is the unit vector representing the accelerometer x-axis, and  $\underline{g}_{now}$  is the gravity vector in the accelerometer frame.




Figure 1: Angle between unit vector along x-axis and gravity vector

We use the definitions of the dot product and cross product between the gravity vector and unit vector along the x-axis to derive the following equations:

$$\cos(\beta) = \frac{\underline{g}_{now} \cdot \hat{x}}{|\underline{g}_{now}| |\hat{x}|} = \frac{g_x}{\sqrt{g_x^2 + g_y^2 + g_z^2}} \quad (1)$$

$$\sin(\beta) = \frac{|\underline{g}_{now} \times \hat{x}|}{|\underline{g}_{now}| |\hat{x}|} = \frac{\sqrt{(g_z)^2 + (g_y)^2}}{\sqrt{g_x^2 + g_y^2 + g_z^2}} \quad (2)$$

As mentioned in the introduction, when  $\sin(\beta)$  or  $\cos(\beta)$  approach 1, their sensitivity to changes in  $\beta$  approaches zero (the derivative tends to zero). Therefore, to maximise accuracy, we use  $\sin(\beta)$  for calculating angles in the ranges  $\beta = 0^\circ - 45^\circ$  and  $\beta = 135^\circ - 180^\circ$  and  $\cos \beta$  for calculating angles in the range  $\beta = 45^\circ - 135^\circ$ . Furthermore, because arcsin does not have a unique solution in the range  $0^\circ$  to  $180^\circ$ , we use the sign of  $\cos(\beta)$  to determine whether the angle given by arcsin lies between  $0^\circ - 45^\circ$  or between  $135^\circ - 180^\circ$ . Therefore, combining the principles described above, the following logical rules can be used to calculate  $\beta$  from  $\sin(\beta)$  and  $\cos(\beta)$ .

If  $\cos(\beta) \geq 0.707107$  (3)

$$\theta = \arcsin(\sin\beta)$$

Else if  $\cos(\beta) \leq -0.707107$

$$\beta = \pi - \arcsin(\sin\beta)$$

Else

$$\beta = \arccos(\cos\beta)$$

End

113

114 The proposed approach overcomes the problem of poor sensitivity to changes in angle when  $\sin(\beta)$   
115 or  $\cos(\beta)$  approach 1 and does not suffer from the singularities seen in approaches which use 2-axis  
116 accelerometer signals and arctan.

117

118 In addition to the approach outlined above we have included an algorithm to reduce the likelihood  
119 of misinterpreting the accelerometer data when the true acceleration becomes significant compared  
120 with  $g$  ( $9.81 \text{ m/s}^2$ ). In cases where the magnitude of the measured accelerometer vector significantly  
121 exceeds  $g$  (i.e. true acceleration is significant compared to  $g$ ) the data point is ignored. This is  
122 achieved by applying a  $g$ -tolerance band ( $9.81 \pm g\text{-tolerance}$ ) and only using good data points that  
123 lie within that band. The FSM controller triggers a state transition when  $n$  good data points (i.e.  
124 within the  $g$ -tolerance band) have exceeded the specified body-segment angle threshold (using  $n$   
125 good points acts as a noise filter). As the  $g$ -tolerance band narrows, it would be expected that a  
126 larger number of bad data points would be ignored by the controller and hence this may lead to a  
127 delay in moving between states. Conversely, if the  $g$ -tolerance band is too wide, errors in angle  
128 estimation would be expected to increase. In this study, we explore the effects of different values  
129 for the  $g$ -tolerance band on angle estimation and the trade-off between accuracy and number of  
130 bad data points ignored.

131

### 132 **Experimental protocol**

133 Following ethical approval (REC ref: 10/H1005/26) two quite different participants, both with upper  
134 limb impairments following stroke, were invited to the lab to participate in the study. The  
135 participants are described in table 1.

136 **Table 1: Participants**

No	Gender	Age	Hemiplegic side	Dominant side	Years since onset	Fugl-Meyer Upper Extremity score (maximum 66)

1	M	81	Left	Right	3 years	29
2	F	42	Right	Right	13 years	37

137

138 Figure 2 shows the experimental setup for the “Drink from a cup” task. Two inertial sensing units or  
139 IMUs (MT9 Xsens bv, NL), each with a cluster of four reflective markers on their upper corners, were  
140 attached to the upper arm and forearm of the subject’s affected limb using self-adherent bandage.  
141 The IMUs’ x-axes were approximately aligned with the long axes of the body segments. A Vicon  
142 motion analysis system (Vicon Motion Systems Ltd, Los Angeles, USA) employing ten cameras was  
143 used to capture the positions of the reflective markers on each IMU at a sampling frequency of 100  
144 Hz. Only acceleration data was captured from the IMUs, at a sampling frequency of 20 Hz, using a  
145 separate laptop. This laptop also ran the Finite State Machine controller that produced the  
146 necessary stimulation profiles via a RehaStim 8-channel stimulator (Hasomed GmbH, Magdeburg,  
147 Germany), and also ran the graphical user interface (GUI) used to set up the FSM controller. A pulse  
148 signal from one of the Xsens system’s analog output channels was fed to an analogue input channel  
149 in the Vicon system to provide synchronization between the Xsens and Vicon systems. This  
150 experimental setup was also used to study another three tasks (see below for details of each task).



151

152 **Figure 2: Experimental setup**

153

154 Each subject was asked to carry out a series of functional tasks assisted by electrical stimulation to  
155 relevant muscles during relevant parts of the movement. The functional tasks were selected by the

therapist at the start of each session as being very difficult or impossible for the user to perform unaided. The four tasks were “Brush coins into the other hand”, “Drink from a cup”, “Place an object onto a shelf” and “Pour water from a bottle to a cup”. For all tasks, the subject sat at a table with their affected hand comfortably placed on the table or on the thigh (for “Place an object into a shelf”) at the starting position. For each repeat of a task, the object(s) to be manipulated was/were placed in a pre-defined starting location.

The therapist used the GUI mentioned earlier to define each task as a sequence of FSM states (corresponding to movement phases), each of which was associated with a stimulation profile<sup>1</sup> for each of the muscles to be stimulated (see figure 3). Threshold values for each muscle were established earlier in the setup process, leaving the therapist to define pulse width target and ramp time for each stimulated muscle in each phase. Pressing a button on the keyboard to leave the neutral phase, the therapist then used the GUI to manually adjust the pulse width target and ramp time for each of the stimulated muscles in phase 1, until the relevant limb motion was achieved. This process was repeated for each movement phase (FSM state). During each attempt, data were recorded from both the Vicon and Xsens systems. Once the therapist was satisfied with the resultant movement, he/she invited the participant to repeat the task until data on between 7 and 10 satisfactory repeats had been captured.

Descriptions of the finite state machines for each of the four functional tasks are as follows:




Figure 3: Finite state diagrams for the following tasks (a) “Brush coins into the other hand”; (b) “Drink from a cup”; (c) Place an object onto a shelf; (d) Pour water from a bottle to a cup. The controller moves between states when the transition condition *Btn* is met (*Btn* = Button press). In each state (represented by a box), the specified muscles are stimulated where: *AD&Tr* = Anterior deltoid and Triceps; *Bi* = Biceps; *Pr* = Pronator; *WE* = Wrist extensors.

<sup>1</sup> A stimulation profile consists of a threshold pulse width, a target pulse width and a ramp time.



180 “Brush coins into the other hand” (figure 3a)

181 The subject was required to reach for coins positioned on the table and brush them back into his/her  
182 other hand. The position of the coins was such that (s)he could only achieve the task with FES  
183 assistance.

184 “Drink from a cup” (figure 3b)

185 The subject was required to reach for a cup, grasp it, lift the cup to the mouth, replace the cup and  
186 release it. The cup was positioned such that the subject could only achieve the task with FES  
187 assistance.

188 “Place an object onto a shelf” (figure 3c)

189 The subject was required to lift his/her forearm towards an object, grasp it, reach forward to put it  
190 on to a shelf and release it. The shelf was located such that (s)he could only achieve the task with  
191 FES assistance.

192  
193 “Pour water from a bottle to a cup” (figure 3d)

194 The subject was required to reach for a bottle, grasp it and pour the water into a cup, replace the  
195 bottle and release it. The position of the cup was such that the subject could only achieve the task  
196 with FES assistance.

197

## 198 ***Data processing***

199

200 The absolute angles from vertical of the two IMU x-axes, based on the accelerometer signals and the  
201 new algorithm described earlier, were recorded directly by the real-time FSM controller. The  
202 coordinates of the reflective markers attached to the IMUs were exported using Visual 3D software  
203 (C-Motion, Inc., Rockville, MD, USA). The Vicon marker data were down-sampled to provide data at  
204 20Hz (frequency of the FSM controller) and synchronized with the IMU data. The calculation of the  
205 angles from vertical of the x-axes of the two IMUs, based on the Vicon data, is described in [32] and  
206 was implemented using Matlab (Mathworks inc. Natick, USA).

207

208 Data were checked post-collection and task repeats were discarded in cases where the marker  
209 visibility was incomplete, or synchronisation between the Xsens and Vicon systems failed.

210

211 To account for small misalignment errors between the marker-derived sensor coordinate frame and  
212 the accelerometer coordinate frame, the first 10 frames of static data were used to artificially  
213 remove the offset. Comparisons between accelerometer and marker-derived angles were drawn  
214 using RMS error and Pearson Correlation coefficients, before and after removing the offset.

215

216 Finally, to investigate the effect of different  $g$ -tolerance bands on the angle estimation, we applied  
217 three different tolerance bands to the data ( $9.81 \pm 0.5 \text{ m/s}^2$ ,  $9.81 \pm 0.3 \text{ m/s}^2$  and  $9.81 \pm$   
218  $0.2 \text{ m/s}^2$ ). Comparisons are presented between maximum error and number of data points lying  
219 outside of the tolerance band.

220

## 221 **Results**

Figure 4 shows example data for forearm and upper arm angles from the vertical, obtained from both reflective marker and accelerometer approaches, for the task “Brushing coins into the other hand” (Left: subject 1; Right: subject 2)




Figure 4: Example data from task “Brush coins into the other hand”

Table 2 compares the angles derived from marker data with accelerometer-derived angles for each subject and each task.

**Table 2: Comparison between marker and accelerometer-derived angles (7-10 trials per subject). Pearson’s correlation ( $r$ ) and RMS error ( $\epsilon$ ) are shown before and after removal of alignment error. (UA = upper arm; FA = forearm)**

**Subject 1:**

	Task 1		Task 2		Task 3		Task 4	
	FA	UA	FA	UA	FA	UA	FA	UA
$r$	0.947	0.986	0.992	0.993	0.943	0.988	0.985	0.994
$\epsilon$ (deg.)	3.12	2.27	2.48	1.62	1.61	1.67	1.63	2.06
$\epsilon$ after removal of alignment error (deg.)	2.91	2.46	1.73	1.60	1.56	1.50	1.72	2.75

**Subject 2:**

	Task 1		Task 2		Task 3		Task 4	
	FA	UA	FA	UA	FA	UA	FA	UA
$r$	0.986	0.995	0.995	0.996	0.985	0.998	0.982	0.990
$\epsilon$ (deg.)	2.82	3.47	1.94	3.24	2.12	3.18	1.91	1.93
$\epsilon$ after removal of alignment error (deg.)	2.33	2.04	2.02	1.42	1.95	2.25	1.66	2.03

Figure 5 shows how maximum errors are reduced, but not eliminated by removing data points which lie outside of the  $g$ -tolerance band.




Figure 5: Errors in forearm “absolute angle from vertical” calculated from an example trial of the “Drink from a cup” task (subject 1) before and after removal of data points using a  $g$ -tolerance  $\pm 0.5 \text{ m/s}^2$ .

Table 3 illustrates the effect of different  $g$ -tolerance bands on the accuracy of angle estimation and number of data points lying outside of the tolerance band.

Table 3: Effect of  $g$ -tolerance (on maximum error ( $\delta$ ) (.deg) and percentage of invalid data ( $p$ ) (%)

Subject 1:

	Tolerance band	Task 1		Task 2		Task 3		Task 4	
		FA	UA	FA	UA	FA	UA	FA	UA
$\delta$	Infinite	12.0 $\pm$ 3.3	7.9 $\pm$ 1.8	12.5 $\pm$ 4.0	8.3 $\pm$ 1.8	8.7 $\pm$ 4.1	8.4 $\pm$ 4.1	7.6 $\pm$ 3.6	5.6 $\pm$ 1.2
	$\pm 0.5 \text{ m/s}^2$	11.1 $\pm$ 2.5	7.3 $\pm$ 1.4	10.1 $\pm$ 2.1	8.1 $\pm$ 1.7	7.2 $\pm$ 2.0	7.9 $\pm$ 4.0	7.4 $\pm$ 3.7	5.5 $\pm$ 1.3
	$\pm 0.3 \text{ m/s}^2$	10.6 $\pm$ 2.6	6.8 $\pm$ 1.6	9.7 $\pm$ 2.4	7.8 $\pm$ 1.5	7.2 $\pm$ 2.0	7.7 $\pm$ 3.9	6.8 $\pm$ 3.8	5.2 $\pm$ 1

	$\pm 0.2 \text{ m/s}^2$	10.6 $\pm$ 2.6	6.5 $\pm$ 1.4	8.1 $\pm$ 2.3	6.8 $\pm$ 1.2	6.1 $\pm$ 2.4	7.3 $\pm$ 4.1	6.6 $\pm$ 3.8	4.9 $\pm$ 0.9
p	Infinite	0	0	0	0	0	0	0	0
	$\pm 0.5 \text{ m/s}^2$	6.5 $\pm$ 2.9	7.1 $\pm$ 3.6	7.5 $\pm$ 2.2	1.1 $\pm$ 0.9	12.7 $\pm$ 3.5	2.1 $\pm$ 1.4	8.0 $\pm$ 3.3	5.1 $\pm$ 2.3
	$\pm 0.3 \text{ m/s}^2$	17.4 $\pm$ 4.5	22.1 $\pm$ 8.7	16.0 $\pm$ 3.6	6.5 $\pm$ 2.1	33.1 $\pm$ 6.8	8.6 $\pm$ 2.6	18.1 $\pm$ 3.7	17.6 $\pm$ 4.2
	$\pm 0.2 \text{ m/s}^2$	28.1 $\pm$ 6.2	34.7 $\pm$ 12.6	30.1 $\pm$ 5.4	15.9 $\pm$ 3.1	51.6 $\pm$ 9.7	19.8 $\pm$ 3.8	32.3 $\pm$ 4.6	30.1 $\pm$ 5.4

## Subject 2:

	Tolerance band	Task 1		Task 2		Task 3		Task 4	
		FA	UA	FA	UA	FA	UA	FA	UA
$\delta$	Infinite	11.3 $\pm$ 3.2	9.5 $\pm$ 1.4	6.5 $\pm$ 2.6	7.2 $\pm$ 0.9	10.27 $\pm$ 4.4	9.43 $\pm$ 2.3	8.16 $\pm$ 1.3	5.5 $\pm$ 0.6
	$\pm 0.5 \text{ m/s}^2$	8.9 $\pm$ 3.7	8.9 $\pm$ 1.8	6.1 $\pm$ 2.8	7.0 $\pm$ 0.8	6.83 $\pm$ 1.1	8.29 $\pm$ 3.1	7.39 $\pm$ 1.8	5.5 $\pm$ 0.6
	$\pm 0.3 \text{ m/s}^2$	8.1 $\pm$ 2.4	8.1 $\pm$ 1.8	4.9 $\pm$ 1.0	6.3 $\pm$ 0.8	6.20 $\pm$ 1.0	7.34 $\pm$ 2.7	7.39 $\pm$ 1.8	5.1 $\pm$ 0.7
	$\pm 0.2 \text{ m/s}^2$	8.1 $\pm$ 2.4	7.3 $\pm$ 1.3	4.6 $\pm$ 0.6	6.1 $\pm$ 1	5.50 $\pm$ 1.0	6.53 $\pm$ 2.0	6.86 $\pm$ 1.7	4.8 $\pm$ 0.8
p	Infinite	0	0	0	0	0	0	0	0
	$\pm 0.5 \text{ m/s}^2$	9.5 $\pm$ 3.0	15.2 $\pm$ 2.8	3.8 $\pm$ 1.3	2.2 $\pm$ 1.2	8.6 $\pm$ 1.5	15.9 $\pm$ 2.9	7.9 $\pm$ 2.4	3.9 $\pm$ 3.1
	$\pm 0.3 \text{ m/s}^2$	14.8 $\pm$ 4.3	35.4 $\pm$ 5.4	12.9 $\pm$ 2.2	30.2 $\pm$ 5.4	20.1 $\pm$ 3.6	33.3 $\pm$ 3.7	19.3 $\pm$ 4.1	27.6 $\pm$ 5.8
	$\pm 0.2 \text{ m/s}^2$	20.3 $\pm$ 4.8	50.8 $\pm$ 7.1	23.6 $\pm$ 2.6	61.0 $\pm$ 4.8	30.7 $\pm$ 5.1	44.7 $\pm$ 5.0	31.9 $\pm$ 6.0	49.4 $\pm$ 6.1

## Discussion and conclusions

This paper has introduced a new method of calculating angle from vertical from a 3-axis accelerometer. The approach avoids the key limitation of methods based on single or dual axis accelerometer signals, namely when the magnitude of the accelerometer signal on the sensitive axis approaches either 9.81 or -9.81, the sensitivity approaches zero and hence the signal to noise ratio becomes very poor. Figure 4 and table 2 illustrate the performance of the method using upper limb mounted IMUs during the performance of a range of typical FES assisted upper limb tasks.

Ignoring readings where the true acceleration is significant in comparison to gravity can remove some unwanted spikes and thereby improve the robustness of angle triggering (Table 3). However, referring to Figures 5, it is clear that not all peaks in error value are removed. This is because only those peaks that alter the magnitude of the measured vector are interpreted as bad readings. Clearly, depending on the direction of the true acceleration (or the equivalent noise from some other source) the magnitude of the measured accelerometer vector may not fall outside the  $g$ -tolerance band. As can be seen in table 3, the tighter the  $g$ -tolerance band the greater the number of bad data points which fall outside and hence the higher the potential delays in transitioning between FSM states using angle triggering. In the worst case, when the  $g$ -tolerance is set at  $\pm 0.2 \text{ m/s}^2$ , 61% of data points fall outside the tolerance band. Based on our findings a  $g$ -tolerance of  $\pm 0.5 \text{ m/s}^2$  appears to be an acceptable value and this was used in subsequent usability trials [33].

The approach shows promise for the application described in the paper. However, the method would not be applicable to limb segments which experience significant accelerations during normal daily activity (e.g. the shank during gait).

- 273 [1] Howlett O, Lannin NA, Ada L, McKinstry C. Functional electrical stimulation improves activity after  
274 stroke: A systematic review with meta-analysis. *Arch Phys Med Rehabil.* 2015.
- 275 [2] Daly JJ, Hogan N, Perepezko EM, Krebs HI, Rogers JM, Goyal KS, et al. Response to upper-limb  
276 robotics and functional neuromuscular stimulation following stroke. *J Rehabil Res Dev.* 2005;42:723-  
277 36.
- 278 [3] Mann GE, Burridge JH, Malone LJ, Strike PW. A pilot study to investigate the effects of electrical  
279 stimulation on recovery of hand function and sensation in subacute stroke patients. *Neuromodul.*  
280 2005;8:193-202.
- 281 [4] Page SJ, Levin L, Hermann V, Dunning K, Levine P. Longer versus shorter daily durations of  
282 electrical stimulation during task-specific practice in moderately impaired stroke. *Arch Phys Med*  
283 *Rehabil.* 2012;93:200-6.
- 284 [5] Tarkka IM, Pitkanen K, Popovic DB, Vanninen R, Kononen M. Functional electrical therapy for  
285 hemiparesis alleviates disability and enhances neuroplasticity. . *Tohoku J Exp Med* 2011;225:71-6.
- 286 [6] Kahn LE, Zygmant ML, Rymer WZ, Reinkensmeyer DJ. Robot-assisted reaching exercise promotes  
287 arm movement recovery in chronic hemiparetic stroke: a randomized controlled pilot study. *J*  
288 *Neuroeng Rehabil.* 2006;3:12.
- 289 [7] Timmermans AA, Seelen HA, Willmann RD, Kingma H. Technology-assisted training of arm-hand  
290 skills in stroke: concepts on reacquisition of motor control and therapist guidelines for rehabilitation  
291 technology design. *J Neuroeng Rehabil.* 2009;6:1.
- 292 [8] Meadmore KL, Hughes AM, Freeman CT, Cai Z, Tong D, Burridge JH, et al. Functional electrical  
293 stimulation mediated by iterative learning control and 3D robotics reduces motor impairment in  
294 chronic stroke. *J Neuroeng Rehabil.* 2012;9:32.
- 295 [9] Thorsen RA, Occhi E, Boccardi S, Ferrarin M. Functional electrical stimulation reinforced tenodesis  
296 effect controlled by myoelectric activity from wrist extensors. *J Rehabil Res Dev.* 2006;43:247-56.
- 297 [10] Thrasher TA, Zivanovic V, McIlroy W, Popovic MR. Rehabilitation of reaching and grasping  
298 function in severe hemiplegic patients using functional electrical stimulation therapy. *Neurorehabil*  
299 *Neural Repair.* 2008;22:706-14.
- 300 [11] Popovic DB, Sinkjaer T, Popovic MB. Electrical stimulation as a means for achieving recovery of  
301 function in stroke patients. *Neurorehabil.* 2009;25:45-58.
- 302 [12] Kim T, Kim S, Lee B. Effects of Action Observational Training Plus Brain-Computer Interface-  
303 Based Functional Electrical Stimulation on Paretic Arm Motor Recovery in Patient with Stroke: A  
304 Randomized Controlled Trial. *Occup Ther Int.* 2015.
- 305 [13] Hara Y, Obayashi S, Tsujiuchi K, Muraoka Y. The effects of electromyography-controlled  
306 functional electrical stimulation on upper extremity function and cortical perfusion in stroke  
307 patients. *Clin Neurophysiol.* 2013;124:2008-15.
- 308 [14] Tresadern P, Thies S, Kenney L, Howard D, Goulermas JY. Artificial Neural Network Prediction  
309 Using Accelerometers to Control Upper Limb FES During Reaching and Grasping Following Stroke.  
310 *IEEE Engineering in Medicine and Biology Society Annual Conference.* 2006;1:2916-9.
- 311 [15] Mann G, Taylor P, Lane R. Accelerometer-triggered electrical stimulation for reach and grasp in  
312 chronic stroke patients: a pilot study. *Neurorehabil Neural Repair.* 2011;25:774-80.
- 313 [16] Tresadern P, Thies S, Kenney LPJ, Howard D, Goulermas JY. Rapid prototyping for functional  
314 electrical stimulation control. *Pervasive Comput.* 2008;7:62-9.
- 315 [17] Tong KY, Mak AF, Ip WY. Command control for functional electrical stimulation hand grasp  
316 systems using miniature accelerometers and gyroscopes. *Med Biol Eng Comput.* 2003;41:710-7.
- 317 [18] Crago PE, Memberg WD, Uzey MK, Keith MK, Kirsch RF, Chapman GJ, et al. An elbow extension  
318 neuroprosthesis for individuals with tetraplegia. *Rehabil Eng.* 1998;6:1-6.
- 319 [19] Van de Dikkenberg N, Meijer OG, Van der Slikke RMA, C VLR, Van Dieën JH, Pijls B, et al.  
320 Measuring functional abilities of patients with knee problems: rationale and construction of the  
321 DynaPort knee test. *Knee Surg, Sports Traumatol, Arthrosc.* 2002;10:204-12.

- [20] Graham RB, Agnew MJ, Stevenson JM. Effectiveness of an on-body lifting aid at reducing low back physical demands during an automotive assembly task: Assessment of EMG response and user acceptability. *Appl Ergon*. 2009;40:936-42.
- [21] Ohnishi K, Kajitani I, Morio T, Takagi T. Multimodal sensor controlled three Degree of Freedom transradial prosthesis. *IEEE International Conference on Rehabilitation Robotics*, 2013. Seattle, Washington: IEEE; 2013. p. 1-6.
- [22] Lugade V, Fortune E, Morrow M, Kaufman K. Validity of Using Tri-Axial Accelerometers to Measure Human Movement - Part I: Posture and Movement Detection. *Med Eng Phys*. 2014;36:169-76.
- [23] Husak M. Model of tilt sensor system. *9th International Conference Electronics, Circuits and Systems: IEEE*; 2002. p. 227-30.
- [24] Bazzarelli M, Durdle NG, Lou E, Raso VJ. A wearable computer for physiotherapeutic scoliosis treatment. *Instrum Measurement*. 2003;52:126-9.
- [25] Baek J, Yun BJ. Posture monitoring system for context awareness in mobile computing. *Instrum Measurement*. 2010;56:1589 - 99.
- [26] Williams AJ. A solid state tilt meter for current meter attitude determination. *Oceans '04 MTTs/IEEE Techno-Ocean '04*. Kobe2004. p. 1395 - 8.
- [27] Rodriguez-Donate C, Morales-Velazquez L, Osornio-Rios RA, Herrera-Ruiz G, Romero-Troncoso RDJ. FPGA-based fused smart sensor for dynamic and vibration parameter extraction in industrial robot links. *Sens*. 2010;10:4114-29.
- [28] Pallejà T, Tresanchez M, Teixidó M, Palacin J. Bioinspired Electronic White Cane Implementation Based on a LIDAR, a Tri-Axial Accelerometer and a Tactile Belt. *Sens*. 2010;10.
- [29] Šipoš M, Pačes P, Roháč J, Nováček P. Analyses of Triaxial Accelerometer Calibration Algorithms. *Sens J*. 2012;12:1157-65.
- [30] Giansanti D, Macellari V, Maccioni G, Cappozzo A. Is it feasible to reconstruct body segment 3-D position and orientation using accelerometric data? *IEEE Trans Biomed Eng*. 2003;50:476-83.
- [31] Madgwick SOH, Harrison AJL, Sharkey PM, Vaidyanathan R, Harwin WS. Measuring motion with kinematically redundant accelerometer arrays: Theory, simulation and implementation. *Mechatron*. 2013;23:518-29.
- [32] Sun M. A functional electrical stimulation (fes) control system for upper limb rehabilitation [Doctoral Thesis]: University of Salford; 2014.
- [33] Smith CL. Usability engineering in the design and evaluation of a functional electrical stimulation system for upper limb rehabilitation [Doctoral Thesis]: University of Salford; 2015.

## Acknowledgements.

This is a summary of independent research funded by the National Institute for Health Research New and Emerging Applications of Technology (NIHR NEAT) Programme (grant ref L030). The views expressed are those of the author(s) and not necessarily those of the NHS, the NIHR or the Department of Health.

**Journal:** MEDICAL ENGINEERING & PHYSICS

**Title of Paper:** Novel methods of using accelerometry for upper limb FES control

### Declarations

The following additional information is required for submission. Please note that failure to respond to these questions/statements will mean your submission will be returned to you. If you have nothing to declare in any of these categories then this should be stated.

#### Conflict of interest

All authors must disclose any financial and personal relationships with other people or organisations that could inappropriately influence (bias) their work. Examples of potential conflicts of interest include employment, consultancies, stock ownership, honoraria, paid expert testimony, patent applications/registrations, and grants or other funding.

#### Ethical Approval

Work on human beings that is submitted to *Medical Engineering & Physics* should comply with the principles laid down in the Declaration of Helsinki; Recommendations guiding physicians in biomedical research involving human subjects. Adopted by the 18th World Medical Assembly, Helsinki, Finland, June 1964, amended by the 29th World Medical Assembly, Tokyo, Japan, October 1975, the 35th World Medical Assembly, Venice, Italy, October 1983, and the 41st World Medical Assembly, Hong Kong, September 1989. You should include information as to whether the work has been approved by the appropriate ethical committees related to the institution(s) in which it was performed and that subjects gave informed consent to the work.

#### Competing Interests

None

#### Please state any sources of funding for your research

National Institute for Health Research New and Emerging Applications of Technology (NIHR NEAT) Programme (grant ref L030)

**DOES YOUR STUDY INVOLVE HUMAN SUBJECTS? Please cross out whichever is not applicable.**

**Yes** ✓

**No**

**If your study involves human subjects you MUST have obtained ethical approval.**

**Please state whether Ethical Approval was given, by whom and the relevant Judgement's reference number**

Yes, we have obtained Ethical Approval.

REC ref: 10/H1005/26

394

395 **This information must also be inserted into your manuscript under the acknowledgements**  
396 **section prior to the References.**

397